

FLIR A325sc

P/N: 48001-1001

Copyright

© 2018, FLIR Systems, Inc.

All rights reserved worldwide. Names and marks appearing herein are either registered trademarks or trademarks of FLIR Systems and/or its subsidiaries. All other trademarks, trade names or company names referenced herein are used for identification only and are the property of their respective owners.

Document identity

Publ. No.: 48001-1001 Release: Commit: 35207 Language: en-US Modified: 2016-04-27 Formatted: 2018-05-24

Website

http://www.flir.com

Customer support

http://support.flir.com

Disclaimer

Specifications subject to change without further notice. Camera models and accessories subject to regional market considerations. License procedures may apply. Products described herein may be subject to US Export Regulations. Please refer to exportquestions@flir.com with any questions.



General description

The FLIR A325sc is an excellent choice for those working in R&D and need high frame rates but for whom 320×240 pixel resolution is sufficient. When using the camera in R&D, it is highly recommended to use the FLIR ResearchIR software from FLIR Systems.

Key features:

- Affordable.
- 16-bit 320×240 pixel images at 60 Hz.
- Start-and-stop recording in FLIR ResearchIR using digital input.
- Lenses: 25° included, 15° and 45° optional.

Typical applications:

• Entry- or mid-level industrial R&D.

Imaging and optical data	
IR resolution	320 × 240 pixels
Thermal sensitivity/NETD	< 0.05°C @ +30°C (+86°F) / 50 mK
Field of view (FOV)	25° × 18.8°
Minimum focus distance	0.4 m (1.31 ft.)
Focal length	18 mm (0.7 in.)
Spatial resolution (IFOV)	1.36 mrad
Lens identification	Automatic
F-number	1.3
Image frequency	60 Hz
Focus	Automatic or manual (built in motor)

Detector data	
Detector type	Focal plane array (FPA), uncooled microbolometer
Spectral range	7.5–13 μm
Detector pitch	25 μm
Detector time constant	Typical 12 ms

Measurement	
Object temperature range	 -20 to +120°C (-4 to +248°F) 0 to +350°C (+32 to +662°F)
Accuracy	±2°C (±3.6°F) or ±2% of reading

1 (14) www.flir.com



FLIR A325sc

P/N: 48001-1001

© 2018, FLIR Systems, Inc. #48001-1001; r. /35207; en-US

Measurement analysis		
Atmospheric transmission correction	Automatic, based on inputs for distance, atmospheric temperature and relative humidity	
Optics transmission correction	Automatic, based on signals from internal sensors	
Emissivity correction	Variable from 0.01 to 1.0	
Reflected apparent temperature correction	Automatic, based on input of reflected temperature	
External optics/windows correction	Automatic, based on input of optics/window transmission and temperature	
Measurement corrections	Global object parameters	
Ethernet		
Ethernet	Control and image	
Ethernet, type	Gigabit Ethernet	
Ethernet, standard	IEEE 802.3	
Ethernet, connector type	RJ-45	
Ethernet, communication	TCP/IP socket-based FLIR proprietary and GenICam protocol	
Ethernet, image streaming	16-bit 320 × 240 pixels @ 60 Hz	
	Signal linearTemperature linearRadiometric	
	GigE Vision and GenlCam compatible	
Ethernet, protocols	TCP, UDP, SNTP, RTSP, RTP, HTTP, ICMP, IGMP, ftp, SMTP, SMB (CIFS), DHCP, MDNS (Bonjour), uPnP	
Digital input/output		
Digital input, purpose	Image tag (start, stop, general), Image flow control, (stream on/off), Input ext. device (programmatically read)	
Digital input	2 opto-isolated, 0–1.5 V = low, 3–25 V = high	
Digital output, purpose	Output to ext. device (programmatically set)	
Digital output	2 opto-isolated, ON = supply (max. 100 mA), OFF = open	
Digital I/O, isolation voltage	500 VRMS	
Digital I/O, supply voltage	6-24 VDC, max. 200 mA	
Digital I/O, connector type	6-pole jackable screw terminal	
Power system		
External power operation	12/24 VDC, 24 W absolute max.	
External power, connector type	2-pole jackable screw terminal	
Voltage	Allowed range 10–30 VDC	
Environmental data		
Operating temperature range	-15°C to +50°C (+5°F to +122°F)	
Storage temperature range	-40°C to +70°C (-40°F to +158°F)	
Humidity (operating and storage)	IEC 60068-2-30/24 h 95% relative humidity +25° C to +40°C (+77°F to +104°F)	

2 (14) www.flir.com

\$FLIR®

FLIR A325sc

P/N: 48001-1001

© 2018, FLIR Systems, Inc. #48001-1001; r. /35207; en-US

Environmental data	
EMC	EN 61000-6-2:2001 (Immunity) EN 61000-6-3:2001 (Emission) FCC 47 CFR Part 15 Class B (Emission)
Encapsulation	IP 40 (IEC 60529)
Shock	25 g (IEC 60068-2-27)
Vibration	2 g (IEC 60068-2-6)

Physical data	
Weight	0.7 kg (1.54 lb.)
Camera size $(L \times W \times H)$	170 × 70 × 70 mm (6.7 × 2.8 × 2.8 in.)
Tripod mounting	UNC 1/4"-20 (on three sides)
Base mounting	$2 \times M4$ thread mounting holes (on three sides)
Housing material	Aluminum

Shipping information	
Packaging, type	Cardboard box
List of contents	Infrared camera with lens Ethernet cable FLIR ResearchIR Max 4 (licence only) Hard transport case Mains cable Power cable, pig-tailed Power supply Printed documentation
Packaging, weight	5.0 kg (11.0 lb.)
Packaging, size	495 × 370 × 192 mm (19.5 × 14.6 × 7.6 in.)
EAN-13	7332558004203
UPC-12	845188004231
Country of origin	Sweden

Supplies & accessories:

- 1196961; IR lens, f = 30 mm, 15° incl. case
- 1196960; IR lens, f = 10 mm, 45° incl. case
- T197215; Close-up 4× (100 μm) incl. case
- T197214; Close-up 2× (50 μm) incl. case
- T197407; IR lens, 76 mm (6°) with case and mounting support for A3xx, A3xxsc
- T197411; IR lens, 4 mm (90°) with case and mounting support for A3xx, A3xxsc
- T197415; Close-up 1× (25 μm) incl. case and mounting support for A3xx, A3xxsc
- T129252; Special temperature range -20 to +700 deg C
- T129253; Special temperature range -20 to +500 deg C
- T129254; High temperature measurement option -20 to +2000 deg C
- T130151; Special temperature range -20 to +2000 deg C
- T130152; Special temperature range +200 to +1200 deg C
- 1910400; Power cord EU
- 1910402; Power cord UK
- 1910401; Power cord US
- T911803; Power supply, 24 VDC, 2 A, 50 W
- T910922; Power supply, incl. multi plugs, for A3xx, A3xxsc, A6xx and A6xxsc
- T951004ACC; Ethernet cable CAT6, 2 m/6.6 ft.
- T911307ACC; Ethernet cable, CAT6, 2 m/6.6 ft, 1 screw connector
- 1910586ACC; Power cable, pigtailed
- T197871ACC; Hard transport case for A3xx/A6xx series
- T197870ACC; Cardboard box for A3xx/A6xx series

3 (14) www.flir.com


\$FLIR


FLIR A325sc


P/N: 48001-1001


© 2018, FLIR Systems, Inc. #48001-1001; r. /35207; en-US


- T198584; FLIR Tools
- T198583; FLIR Tools+ (download card incl. license key)
- T198697; FLIR ResearchIR Max + HSDR 4 (hardware sec. dev.)
- T199014; FLIR ResearchIR Max + HSDR 4 (printed license key)
- T199044; FLIR ResearchIR Max + HSDR 4 Upgrade (printed license key)
- T198696; FLIR ResearchIR Max 4 (hardware sec. dev.)
- T199013; FLIR ResearchIR Max 4 (printed license key)
- T199043; FLIR ResearchIR Max 4 Upgrade (printed license key)
- T198731; FLIR ResearchIR Standard 4 (hardware sec. dev.)
- T199012; FLIR ResearchIR Standard 4 (printed license key)
- T199042; FLIR ResearchIR Standard 4 Upgrade (printed license key)
- . T199233; FLIR Atlas SDK for .NET
- T199234; FLIR Atlas SDK for MATLAB
- T198567; ThermoVision™ System Developers Kit Ver. 2.6
- T198566; ThermoVision™ LabVIEW® Digital Toolkit Ver. 3.3
- INST-EW-0150; Extended Warranty 1 Year for A3xx, B/T440, B/T450sc, T460
- INST-EWGM-0155; Extended Premier Warranty 1 Year for A3xx, T540, T600/bx, T610
- INST-GM-0150; Calibration incl General Maintenance for A3xx, T540, T6xx




















 $I_{MAX} = 100 \text{ mA}$ Low = 0-1.5 VHigh = 3-25 V ${\bf R}_{\mathsf{LOAD}}$ ⊃[™] 7ੂ 6-24 V U_{SUPPLY} 4 • OUT 2 5 Z EMI-FILTERING SIGNAL CONDITIONING EMI-FILTERING SIGNAL CONDITIONING SECONDARY EMI-FILTERING EMI-FILTERING EMI-FILTERING I/O Ground Camera Ground PRIMARY Digital I/O Control

Digital I/O connection diagrams for FLIR A3xx/A6xx series



April 24, 2017 Täby, Sweden

AQ320234

CE Declaration of Conformity - EU Declaration of Conformity

Product: FLIR A3XX -series including A3XXSC

Name and address of the manufacturer: FLIR Systems AB PO Box 7376 SE-187 15 Täby, Sweden

This declaration of conformity is issued under the sole responsibility of the manufacturer.

The object of the declaration: FLIR A3XX -series including A3XXSC.

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation:

Directives:

Directive

2014/30/EU

Electromagnetic Compability

Directive Directive

2014/35/EU 2012/19/EU Low Voltage Directive (Power Supply)

Waste electrical and electric equipment

Standards:

Emission:

EN 61000-6-3:2006

Electromagnetic Compability

Generic standards – Emission

Immunity:

EN 61000-6-2:2005

Electromagnetic Compability

Generic standards – Immunity

Safety (Power supply):

EN 60950-1

Information technology equipment

FLIR Systems ABQuality Assurance

Lea Dabiri

Quality Manager